

# Measurement Equipment Report

## Multimeters



ID No: ~~004~~ (1)

Brand: FLUKE

Model 75

Serial No: 74811094

Test Frequency 12 MONTH.

### Closed Circuit Potential Test

Test Date	Master	This Meter	Dev.	OK?
9/12/13	-1.062	-1.058	0.004	YES

### Voltage Test 2.77V

Test Date	Master	This Meter	Dev.	OK?
9/12/13	2.778	2.784	0.006	YES

### Voltage Test 12V

Test Date	Master	This Meter	Dev.	OK?
9/12/13	12.026	12.084	0.058	YES

Tested By: R. NORTHAM Date: 9/12/2013