

Measurement Equipment Report

Multimeters



ID No: No 1

Brand: FLUKE

Model 75

Serial No: 19811099

Test Frequency 12 MONTH

Closed Circuit Potential Test

Test Date	Master	This Meter	Dev.	OK?
28/11/14	-1.062	-1.063	0.001	YES

Voltage Test 2.77V

Test Date	Master	This Meter	Dev.	OK?
28/11/14	2.80	2.85	0.05	YES

Voltage Test 12V

Test Date	Master	This Meter	Dev.	OK?
28/11/14	11.80	11.87	0.007	YES

Tested By: R N.

Date: 28/11/2014