

Measurement Equipment Report

Multimeters



ID No: 21

Brand: RISHABH

Model RISH MULTI 145 MASTER.

Serial No: 090891

Test Frequency MASTER 48 MONTHS.

Closed Circuit Potential Test

Test Date	Master	This Meter	Dev.	OK?
28/11/2014		-1.062	NA	YES.

Voltage Test 2.77V

Test Date	Master	This Meter	Dev.	OK?
28/11/2014		2.80	NA	YES

Voltage Test 12V

Test Date	Master	This Meter	Dev.	OK?
28/11/2014		11.80	NA	YES

Tested By: RN.

Date: 28/11/2014.