

Measurement Equipment Report

Multimeters



ID No: 22

Brand: DICK SMITH

Model Q1419

Serial No: 180557275

Test Frequency 12 MONTH

Closed Circuit Potential Test

Test Date	Master	This Meter	Dev.	OK?
9/12/13	-1.062	-1.06	0.002	YES

Voltage Test 2.77V

Test Date	Master	This Meter	Dev.	OK?
9/12/13	2.748	2.80	0.022	YES

Voltage Test 12V

Test Date	Master	This Meter	Dev.	OK?
9/12/13	12.026	12.10	0.074	YES

Tested By: R. NORTHBY Date: 9/12/2013