

ID No.	Item	Brand	Model	Serial No.	Location	Test Frequency							
22	Multimeter	Dick Smith	Q-1419	180557275	Laboratory	12 monthly							
Test Date	Closed Circuit Potential Test				Test Date	Voltage Test (12V)				Voltage Test (2.77V)			
25SEP09	Master	This Meter	Dev.	OK?	20OCT09	Master	This Meter	Dev.	OK?	Master	This Meter	Dev.	OK?
	-899mV	-882mV	17mV	OK		12.22V	12.23V	0.01V	OK	2.77V	2.78V	0.01V	OK
Test Date	Closed Circuit Potential Test				Test Date	Voltage Test (12V)				Voltage Test (2.77V)			
17NOV 2010	Master	This Meter	Dev.	OK?	17NOV 2010	Master	This Meter	Dev.	OK?	Master	This Meter	Dev.	OK?
	101.5	97.3	4.2	Y		12.05	12.07	0.001	Y	2.996	2.99	0.006	Y
Test Date	Closed Circuit Potential Test				Test Date	Voltage Test (12V)				Voltage Test (2.77V)			
	Master	This Meter	Dev.	OK?		Master	This Meter	Dev.	OK?	Master	This Meter	Dev.	OK?
Test Date	Closed Circuit Potential Test				Test Date	Voltage Test (12V)				Voltage Test (2.77V)			
	Master	This Meter	Dev.	OK?		Master	This Meter	Dev.	OK?	Master	This Meter	Dev.	OK?