

Measurement Equipment Report

Multimeters



ID No: 22

Brand: DICK SMITH

Model Q1419.

Serial No: 180557275

Test Frequency 12 MONTH

Closed Circuit Potential Test

Test Date	Master	This Meter	Dev.	OK?
28/11/14	-1.062	-1.05	0.012	YES

Voltage Test 2.77V

Test Date	Master	This Meter	Dev.	OK?
28/11/14	2.80	2.90	0.10	YES

Voltage Test 12V

Test Date	Master	This Meter	Dev.	OK?
28/11/14	11.80	11.90	0.10	YES

Tested By: RN

Date: 28/11/2014.