

# Measurement Equipment Report

## Multimeters



ID No: 23

Brand: DIGITECH

Model QM1535

Serial No: 795773

Test Frequency 12 MONTHLY

### Closed Circuit Potential Test

Test Date	Master	This Meter	Dev.	OK?
9/12/13	-1.062	-1.056	0.006	YES

### Voltage Test 2.77V

Test Date	Master	This Meter	Dev.	OK?
9/12/13	2.778	2.788	0.010	YES

### Voltage Test 12V

Test Date	Master	This Meter	Dev.	OK?
9/12/13	12.026	12.096	0.070	YES

Tested By: R. NORTHBY Date: 9/12/2013