

ID No. <b>23</b>	Item <b>Multimeter</b>	Brand <b>Digitech</b>	Model <b>QM</b>	Serial No. <b>0795773</b>	Location <b>Laboratory</b>	Test Frequency <b>12 monthly</b>							
Test Date	Closed Circuit Potential Test				Test Date	Voltage Test (12V)				Voltage Test (2.77V)			
	Master	This Meter	Dev.	OK?	20OCT09	Master	This Meter	Dev.	OK?	Master	This Meter	Dev.	OK?
						12.22V	12.25V	0.03V	OK	2.77V	2.79V	0.02V	OK
Test Date	Closed Circuit Potential Test				Test Date	Voltage Test (12V)				Voltage Test (2.77V)			
	Master	This Meter	Dev.	OK?		Master	This Meter	Dev.	OK?	Master	This Meter	Dev.	OK?
Test Date	Closed Circuit Potential Test				Test Date	Voltage Test (12V)				Voltage Test (2.77V)			
	Master	This Meter	Dev.	OK?		Master	This Meter	Dev.	OK?	Master	This Meter	Dev.	OK?
Test Date	Closed Circuit Potential Test				Test Date	Voltage Test (12V)				Voltage Test (2.77V)			
	Master	This Meter	Dev.	OK?		Master	This Meter	Dev.	OK?	Master	This Meter	Dev.	OK?