

Measurement Equipment Report

Multimeters



ID No: 23

Brand: DIGITREX

Model QM1535

Serial No: 795473

Test Frequency 12 MONTH

Closed Circuit Potential Test

Test Date	Master	This Meter	Dev.	OK?
28/11/14	-1.062	-1.065	0.003	YES

Voltage Test 2.77V

Test Date	Master	This Meter	Dev.	OK?
28/11/14	2.80	2.82	0.02	YES

Voltage Test 12V

Test Date	Master	This Meter	Dev.	OK?
28/11/14	11.80	11.85	0.05	YES

Tested By: RMW

Date: 28/11/2014