

ID No.	Item	Brand	Model	Serial No.	Location	Test Frequency							
23	Multimeter	Digitech	QM	0795773	Laboratory	12 monthly							
Test Date	Closed Circuit Potential Test				Test Date	Voltage Test (12V)				Voltage Test (2.77V)			
	Master	This Meter	Dev.	OK?	20OCT09	Master	This Meter	Dev.	OK?	Master	This Meter	Dev.	OK?
						12.22V	12.25V	0.03V	OK	2.77V	2.79V	0.02V	OK
Test Date	Closed Circuit Potential Test				Test Date	Voltage Test (12V)				Voltage Test (2.77V)			
17 NOV 2010	Master	This Meter	Dev.	OK?	17 NOV 2010	Master	This Meter	Dev.	OK?	Master	This Meter	Dev.	OK?
	101.5	101.2	0.3	Y		12.05	12.05	NIL	Y	2.996	3.002	0.6	Y
Test Date	Closed Circuit Potential Test				Test Date	Voltage Test (12V)				Voltage Test (2.77V)			
31 JAN 2012	Master	This Meter	Dev.	OK?		Master	This Meter	Dev.	OK?	Master	This Meter	Dev.	OK?
	1.087	1.083	0.003	Y		11.886	11.778	-0.8	Y	2.776	2.768	0.008	Y
Test Date	Closed Circuit Potential Test				Test Date	Voltage Test (12V)				Voltage Test (2.77V)			
	Master	This Meter	Dev.	OK?		Master	This Meter	Dev.	OK?	Master	This Meter	Dev.	OK?