


Certificate of Verification

This Ionode electrode model IH 10 & serial number 1404 - 127353
was measured against a Saturated Calomel Reference Electrode (SCE) SN 11195 - A07
Test undertaken at 25 °C in an un-stirred aqueous solution of saturated potassium chloride with a GLP compliant pH / mV meter



Measured potential (mV) against SCE..... -45.0 mV

Tested by: Victor van

Date tested: 1/05 / 2014

Comments:

ionode
GOOD CHEMISTRY



Authorized signatory

[Handwritten Signature]

Dated

01/05 / 2014