

Measuretron (Qld) Pty. Limited

A.C.N 010 921 007 A.B.N 23 010 921 007

192 BANCROFT TERRACE DECEPTION BAY QLD 4508

T/A NORTHSIDE / UNIVERSAL SCALES

TEL: 07 3204-0084 FAX: 07 3204-0092

CALIBRATION REPORT

CUSTOMER:

Cathodic Diecasting
8 Endeavour Drive,
KUNDA PARK QLD 4556

EQUIPMENT:

MAKE Nuveim/Randa NO.
MODEL 1H1949/R5000/N100184/3139434
CAPACITY 3000kg x 1kg
LOCATION
TRADE APPROVED YES NO

CONTACT: Russell Northey

TEST RESULTS

LOAD kg g	AS FOUND kg g	FINAL TEST kg g
0	0	0
500	500	500
1000	1000	1000
1500	1500	1500

COMMENTS:

ALL TESTING CARRIED OUT USING STANDARD MASSES COVERED BY ONE OR MORE OF THE FOLLOWING CERTIFICATES ISSUED TO MEASURETRON BY NATIONAL MEASUREMENT INSTITUTE, BRISBANE REGIONAL OFFICE, MEASUREMENT STANDARDS LABORATORY, GEEBUNG QUEENSLAND. CERTIFICATE NUMBERS: - RN101575, RN111646, RN111647, RN111648

SIGNATURE:

DATE: 01.06.2012

Jeff Kenyon

RE-TEST DUE: May 2013