

Measuretron (Qld) Pty. Limited

A.C.N 010 921 007 A.B.N 23 010 921 007

192 BANCROFT TERRACE DECEPTION BAY QLD 4508

T/A NORTHSIDE / UNIVERSAL SCALES

TEL: 07 3204-0084 FAX: 07 3204-0092

CALIBRATION REPORT

CUSTOMER:

CATHODIC DIECASTING
8 ENDEAVOUR DRIVE
KUNDA PARK QLD 4556
CONTACT: RUSSELL NORTHEY

EQUIPMENT:

MAKE Nissei/Ribson ID NO.
MODEL in 1949/R5000 S/N 100184/3139434
CAPACITY 3000 kg x 1kg
LOCATION _____
TRADE APPROVED YES NO


TEST RESULTS

LOAD kg g	AS FOUND kg g	FINAL TEST kg g
0	0	0
500	500	500
1000	1000	1000
2000	2000	2000
3000	3000	3000

COMMENTS:

ALL TESTING CARRIED OUT USING STANDARD MASSES COVERED BY ONE OR MORE OF THE FOLLOWING CERTIFICATES ISSUED TO MEASURETRON BY NATIONAL MEASUREMENT INSTITUTE, BRISBANE REGIONAL OFFICE, MEASUREMENT STANDARDS LABORATORY, GEEBUNG QUEENSLAND. CERTIFICATE NUMBERS: - RN101532, RN101574, RN101575, RN101576

SIGNATURE:



DATE: 30.5.2011

RE-TEST DUE: MAY 2012