

Measuretron (Qld) Pty. Limited

A.C.N 010 921 007 A.B.N 23 010 921 007

192 BANCROFT TERRACE DECEPTION BAY QLD 4508

T/A NORTHSIDE / UNIVERSAL SCALES

TEL: 07 3204-0084 FAX: 07 3204-0092

CALIBRATION REPORT

CUSTOMER:

Cathodic Anodes Aust.

8 Endeavour Drive,

KUNDA PARK QLD 4556

CONTACT: Russell Northey

EQUIPMENT:

MAKE ISHIDA ID NO. 4

MODEL MB-300E S/N 05340047

CAPACITY 3000g x 0.1g

LOCATION _____

TRADE APPROVED YES NO

TEST RESULTS

LOAD kg g	AS FOUND kg g	FINAL TEST kg g
0	0	0
200	200.0	200.0
500	500.0	500.0
1000	1000.0	1000.0
2000	2000.0	2000.0
3000	3000.0	3000.0

COMMENTS:

ALL TESTING CARRIED OUT USING STANDARD MASSES COVERED BY ONE OR MORE OF THE FOLLOWING CERTIFICATES ISSUED TO MEASURETRON BY NATIONAL MEASUREMENT INSTITUTE, BRISBANE REGIONAL OFFICE, MEASUREMENT STANDARDS LABORATORY, GEEBUNG QUEENSLAND. CERTIFICATE NUMBERS: - RN121908, CN031-441, RN132339 and RN140256.

SIGNATURE:

DATE: 23.07.2014

Jeff Kenyon

RE-TEST DUE: June 2015