

Measuretron (Qld) Pty. Limited

A.C.N 010 921 007 A.B.N 23 010 921 007

192 BANCROFT TERRACE DECEPTION BAY QLD 4508

T/A NORTHSIDE / UNIVERSAL SCALES

TEL: 07 3204-0084 FAX: 07 3204-0092

CALIBRATION REPORT

CUSTOMER:

Cathodic Diecasting

8 Endeavour Drive,

KUNDA PARK QLD 4556

CONTACT: Russell Northey

EQUIPMENT:

MAKE Newey ID NO. 12

MODEL JAD 525 S/N 706187

CAPACITY 15kg x 2g

LOCATION _____

TRADE APPROVED YES NO

TEST RESULTS

| LOAD kg g | AS FOUND kg g | FINAL TEST kg g |
|--------------|------------------|--------------------|
| 0 | 0 | 0 |
| 1 | 1 | 1 |
| 2 | 2 | 2 |
| 5 | 5 | 5 |
| 8 | 7.998 | 8 |
| 10 | 9.996 | 10 |
| 15 | 14.994 | 15 |

COMMENTS:

ALL TESTING CARRIED OUT USING STANDARD MASSES COVERED BY ONE OR MORE OF THE FOLLOWING CERTIFICATES ISSUED TO MEASURETRON BY NATIONAL MEASUREMENT INSTITUTE, BRISBANE REGIONAL OFFICE, MEASUREMENT STANDARDS LABORATORY, GEEBUNG QUEENSLAND. CERTIFICATE NUMBERS: - RN121908, RN121992, CN031-441, RN130043.

SIGNATURE:

DATE: 04.06.2013

Jeff Kenyon

RE-TEST DUE: May 2014