

Measurement Equipment Report

Multimeters



ID No: No 1

Brand: FLUKE

Model 75

Serial No: 79811094

Test Frequency 12 MONTH.

Closed Circuit Potential Test

Test Date	Master	This Meter	Dev.	OK?
11/11/16	-1.063	-1.061	0.002	YES

Voltage Test 2.77V

Test Date	Master	This Meter	Dev.	OK?
11/11/16	2.64	2.64	—	YES

Voltage Test 12V

Test Date	Master	This Meter	Dev.	OK?
11/11/16	14.8	14.8	—	YES

Tested By: R.N.

Date: 11/11/2016