

Measurement Equipment Report

Multimeters



ID No: No 1

Brand: FLUKE.

Model 75.

Serial No: 74811094

Test Frequency 12 MONTH

Closed Circuit Potential Test

Test Date	Master	This Meter	Dev.	OK?
17/11/15	-1.072	-1.071	-0.001	YES

Voltage Test 2.77V

Test Date	Master	This Meter	Dev.	OK?
17/11/15	2.60	2.59.	-01	YES

Voltage Test 12V

Test Date	Master	This Meter	Dev.	OK?
17/11/15	15.3	15.3	—	YES

Tested By: RN.

Date: 17/11/2015.