

Measuretron (Qld) Pty. Limited

A.C.N 010 921 007 A.B.N 23 010 921 007

192 BANCROFT TERRACE DECEPTION BAY QLD 4508

T/A NORTHSIDE / UNIVERSAL SCALES

TEL: 07 3204-0084 FAX: 07 3204-0092

CALIBRATION REPORT

CUSTOMER:

Cathodic Anodes Aust.
8 Endeavour Drive,
KUNDA PARK QLD 4556

EQUIPMENT:

MAKE ISHIDA ID NO. 4
MODEL MB-300E S/N 05340047
CAPACITY 3000g x 0.1g.
LOCATION _____
TRADE APPROVED YES NO

CONTACT: Russell Northey

TEST RESULTS

| LOAD kg g | AS FOUND kg g | FINAL TEST kg g |
|-------------------------|-----------------------------|-------------------------------|
| 0 | 0.0 | 0.0 |
| 200 | 200.0 | 200.0 |
| 500 | 500.0 | 500.0 |
| 1000 | 1000.0 | 1000.0 |
| 2000 | 2000.0 | 2000.0 |
| 3000 | 3000.0 | 3000.0 |
| | | |

COMMENTS:

ALL TESTING CARRIED OUT USING STANDARD MASSES COVERED BY ONE OR MORE OF THE FOLLOWING CERTIFICATES ISSUED TO MEASURETRON BY NATIONAL MEASUREMENT INSTITUTE, BRISBANE REGIONAL OFFICE, MEASUREMENT STANDARDS LABORATORY, GEEBUNG QUEENSLAND. CERTIFICATE NUMBERS: - RN151239, RN151924, RN161940 and CN031-651.

SIGNATURE:

DATE: 25/10/2016

Jeff Kenyon

RE-TEST DUE: October 2017