

SCALE VERIFICATION / CALIBRATION REPORT

Purpose:		<input type="checkbox"/> Trade Verification <input checked="" type="checkbox"/> Non-Trade Calibration <input type="checkbox"/> Other			SI Unit: kg <input checked="" type="checkbox"/> g		
Customer name:		Cathodic Diecasting Qld Pty Ltd					
Customer address:		8 Endeavour Drive, KUNDA PARK QLD 4556					
Instrument location (if different from address above)							
Contact:	Jody	Job No:	1165274	Cust. Ref No:	05		
INSTRUMENT DETAILS:							
Make:	Nagata	Class:	NFTU	Serial No:	14407		
Model:	HB-03	Capacity:	2000 kg	NSC/NMI No:	----		
Type:	Crane Scale	Graduation (e):	2/1/0.5 kg	Old Mark:	----		
VISUAL INSPECTION:							
Comply with NMI Approval: <small>(Trade approved, if applicable)</small>		<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A		Comply with NITP 6.1-6.4: <small>(Visual inspection requirements - Clause 3)</small>			
			<input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A				
REPEATABILITY TEST RESULTS:				ECCENTRICITY TEST RESULTS:			
1.	1000.0	3.	1000.0	----	----	----	
2.	1000.0	Max Diff =	0	----	----	----	
Test Load =	1000 kg	<input checked="" type="checkbox"/> Pass <input type="checkbox"/> Fail		Test Load =	N/A <input type="checkbox"/> Pass <input type="checkbox"/> Fail		
WEIGHING TEST		Initial Weighing Results			Weighing Results after adjustment		
Applied Load Min to Max	MPE +/-	↑ Reading	↓ Reading	P/F	↑ Reading	↓ Reading	P/F
500	----	502.0	/		500.0	500.0	P
1000	----	1004.0			1000.0	1000.0	P
1500	----	1506.0			1499.5	1499.5	P
2000	----	2008.0			1999.5	1999.5	P
Over-range blanking:		<input checked="" type="checkbox"/> Pass <input type="checkbox"/> Fail		Weighing performance:			<input checked="" type="checkbox"/> Pass <input type="checkbox"/> Fail
OTHER PERFORMANCE TEST RESULTS (as applicable):							
Accuracy, Zero indication: <small>(Trade, Initial verification)</small>		<input type="checkbox"/> Pass <input type="checkbox"/> Fail <input checked="" type="checkbox"/> N/A		Sensitivity: <small>(Trade, Non-self indicating instrument only)</small>			<input type="checkbox"/> Pass <input type="checkbox"/> Fail <input checked="" type="checkbox"/> N/A
Accuracy, Tare setting: <small>(Trade, Initial verification)</small>		<input type="checkbox"/> Pass <input type="checkbox"/> Fail <input checked="" type="checkbox"/> N/A		Price Indication:			<input type="checkbox"/> Pass <input type="checkbox"/> Fail <input checked="" type="checkbox"/> N/A
Discrimination: <small>(Trade)</small>		<input type="checkbox"/> Pass <input type="checkbox"/> Fail <input checked="" type="checkbox"/> N/A		Printed information: <small>(Labels / Transaction records)</small>			<input type="checkbox"/> Pass <input type="checkbox"/> Fail <input checked="" type="checkbox"/> N/A
Details of reference masses used		500kg - WS 2005, WS 2008, WS 2009, WS 2010					
Comments & Overall result		ADJUSTED AND CALIBRATED.					
		<input checked="" type="checkbox"/> Pass <input type="checkbox"/> Fail					
Tested by:		B. ELPHINSTONE		Test Date:	6.8.2015		
Name of signatory:		_____		Date Next Due:	AUG. 2016		
Position of signatory:		_____					
Signature		B Elphinstone		Report Date	_____		