

Measuretron (Qld) Pty. Limited

A.C.N 010 921 007 A.B.N 23 010 921 007

192 BANCROFT TERRACE DECEPTION BAY QLD 4508

T/A NORTHSIDE / UNIVERSAL SCALES

TEL: 07 3204-0084 FAX: 07 3204-0092

CALIBRATION REPORT

CUSTOMER:

Cathodic Anodes Aust.
8 Endeavour Drive,
KUNDA PARK QLD 4556

EQUIPMENT:

MAKE Newtech/Randall NO. 6
MODEL 141949/R5000S/N100184/3139434
CAPACITY 3000kg x 1kg
LOCATION _____
TRADE APPROVED YES NO

CONTACT: Russell Northey

TEST RESULTS

| LOAD kg g | AS FOUND kg g | FINAL TEST kg g |
|--------------|------------------|--------------------|
| 0 | | 0 |
| 500 | | 500 |
| 1000 | SEE COMMENTS | 1000 |
| | | |
| | | |
| | | |
| | | |

COMMENTS: SCALE NEED REPLACEMENT POWER SUPPLY
FITTING PRIOR TO TESTING.

ALL TESTING CARRIED OUT USING STANDARD MASSES COVERED BY ONE OR MORE OF THE FOLLOWING CERTIFICATES ISSUED TO MEASURETRON BY NATIONAL MEASUREMENT INSTITUTE, BRISBANE REGIONAL OFFICE, MEASUREMENT STANDARDS LABORATORY, GEEBUNG QUEENSLAND. CERTIFICATE NUMBERS: - RN121908, CN031-441, RN132339 and RN140256.

SIGNATURE:

DATE: 10.04.2014

Jeff Kenyon

RE-TEST DUE: May 2015