

Measuretron (Qld) Pty. Limited

A.C.N 010 921 007 A.B.N 23 010 921 007

192 BANCROFT TERRACE DECEPTION BAY QLD 4508

T/A NORTHSIDE / UNIVERSAL SCALES

TEL: 07 3204-0084 FAX: 07 3204-0092

CALIBRATION REPORT

CUSTOMER:

Cathodic Anodes Aust.

8 Endeavour Drive,

KUNDA PARK QLD 4556

EQUIPMENT:

MAKE RINSTEAM/Ancal ID NO. 07.

MODEL R321 S/N -

CAPACITY 250kg x 0.1kg

LOCATION J

TRADE APPROVED YES NO

CONTACT: Russell Northey

TEST RESULTS

LOAD kg g	AS FOUND kg g	FINAL TEST kg g
0	0	0
20	20	20
40	39.99	40
80	79.98	80
120	119.96	120
160	159.94	160
200	199.91	200

COMMENTS:

ALL TESTING CARRIED OUT USING STANDARD MASSES COVERED BY ONE OR MORE OF THE FOLLOWING CERTIFICATES ISSUED TO MEASURETRON BY NATIONAL MEASUREMENT INSTITUTE, BRISBANE REGIONAL OFFICE, MEASUREMENT STANDARDS LABORATORY, GEEBUNG QUEENSLAND. CERTIFICATE NUMBERS: - RN141872, RN151239, RN151924 and CN031-651.

SIGNATURE:

DATE: 30.09.2015

Jeff Kenyon

RE-TEST DUE: September 2016