

# Measuretron (Qld) Pty. Limited

A.C.N 010 921 007 A.B.N 23 010 921 007

192 BANCROFT TERRACE DECEPTION BAY QLD 4508

T/A NORTHSIDE / UNIVERSAL SCALES

TEL: 07 3204-0084 FAX: 07 3204-0092

## CALIBRATION REPORT

### CUSTOMER:

Cathodic Anodes Aust.  
8 Endeavour Drive,  
KUNDA PARK QLD 4556

### EQUIPMENT:

MAKE Rydstrom ID NO. 07  
MODEL R324 S/N \_\_\_\_\_  
CAPACITY 250kg x 0.1kg  
LOCATION \_\_\_\_\_  
TRADE APPROVED YES NO

CONTACT: Russell Northey

## TEST RESULTS

LOAD kg g	AS FOUND kg g	FINAL TEST kg g
0	0	0
20	19.9	20
40	39.8	40
80		80
120		120

### COMMENTS:

ALL TESTING CARRIED OUT USING STANDARD MASSES COVERED BY ONE OR MORE OF THE FOLLOWING CERTIFICATES ISSUED TO MEASURETRON BY NATIONAL MEASUREMENT INSTITUTE, BRISBANE REGIONAL OFFICE, MEASUREMENT STANDARDS LABORATORY, GEEBUNG QUEENSLAND. CERTIFICATE NUMBERS: - RN151239, RN151924, RN161940 and CN031-651.

SIGNATURE:

DATE: 25/10/2016

*Jeff Kenyon*

RE-TEST DUE: October 2017