

Measuretron (Qld) Pty. Limited

A.C.N 010 921 007 A.B.N 23 010 921 007

192 BANCROFT TERRACE DECEPTION BAY QLD 4508

T/A NORTHSIDE / UNIVERSAL SCALES

TEL: 07 3204-0084 FAX: 07 3204-0092

CALIBRATION REPORT

CUSTOMER:

Cathodic Anodes Aust.
8 Endeavour Drive,
KUNDA PARK QLD 4556

EQUIPMENT:

MAKE NUWEIGH ID NO. 8
MODEL JAC330 S/N E02407-0161
CAPACITY 600kg x 50g
LOCATION _____
TRADE APPROVED YES NO

CONTACT: Russell Northey

TEST RESULTS

LOAD kg g	AS FOUND kg g	FINAL TEST kg g
0		0
50		50
100	SEE COMMENTS	100
200		200
500		500

COMMENTS:

COMPLETE RE-TEST & CALIBRATION
DUE TO REPLACEMENT LOAD CELL BEING FITTED.

ALL TESTING CARRIED OUT USING STANDARD MASSES COVERED BY ONE OR MORE OF THE FOLLOWING CERTIFICATES ISSUED TO MEASURETRON BY NATIONAL MEASUREMENT INSTITUTE, BRISBANE REGIONAL OFFICE, MEASUREMENT STANDARDS LABORATORY, GEEBUNG QUEENSLAND. CERTIFICATE NUMBERS: - RN121908, CN031-441, RN132339 and RN140256.

SIGNATURE:

DATE: 10.04.2014

Jeff Kenyon

RE-TEST DUE: May 2015