

# Measuretron (Qld) Pty. Limited

A.C.N 010 921 007 A.B.N 23 010 921 007

192 BANCROFT TERRACE DECEPTION BAY QLD 4508

T/A NORTHSIDE / UNIVERSAL SCALES

TEL: 07 3204-0084 FAX: 07 3204-0092

## CALIBRATION REPORT

### CUSTOMER:

Cathodic Anodes Aust.

8 Endeavour Drive,

KUNDA PARK QLD 4556

### EQUIPMENT:

MAKE Mercury ID NO. 09

MODEL 522 PTL S/N K 1012

CAPACITY 2500kg x 5kg

LOCATION \_\_\_\_\_

TRADE APPROVED YES NO

CONTACT: Russell Northey

## TEST RESULTS

LOAD kg g	AS FOUND kg g	FINAL TEST kg g
0	0	0
200	200	200
500	500	500
1000	1000	1000

### COMMENTS:

ALL TESTING CARRIED OUT USING STANDARD MASSES COVERED BY ONE OR MORE OF THE FOLLOWING CERTIFICATES ISSUED TO MEASURETRON BY NATIONAL MEASUREMENT INSTITUTE, BRISBANE REGIONAL OFFICE, MEASUREMENT STANDARDS LABORATORY, GEEBUNG QUEENSLAND. CERTIFICATE NUMBERS: - RN121908, CN031-441, RN132339 and RN140256.

SIGNATURE:

DATE: 10.04.2014

*Jeff Kenyon*

RE-TEST DUE: May 2015