

Measuretron (Qld) Pty. Limited

A.C.N 010 921 007 A.B.N 23 010 921 007

192 BANCROFT TERRACE DECEPTION BAY QLD 4508

T/A NORTHSIDE / UNIVERSAL SCALES

TEL: 07 3204-0084 FAX: 07 3204-0092

CALIBRATION REPORT

CUSTOMER:

Cathodic Anodes Aust.

8 Endeavour Drive,

KUNDA PARK QLD 4556

CONTACT: Russell Northey

EQUIPMENT:

MAKE NUMEICH ID NO. 12

MODEL JAD 525 S/N 706187

CAPACITY 15kg x 5g.

LOCATION _____

TRADE APPROVED YES NO

TEST RESULTS

LOAD kg g	AS FOUND kg g	FINAL TEST kg g
0	0	0
1	1	1
2	2	2
5	5	5
8	8	8
10	10	10
15	15	15

COMMENTS:

ALL TESTING CARRIED OUT USING STANDARD MASSES COVERED BY ONE OR MORE OF THE FOLLOWING CERTIFICATES ISSUED TO MEASURETRON BY NATIONAL MEASUREMENT INSTITUTE, BRISBANE REGIONAL OFFICE, MEASUREMENT STANDARDS LABORATORY, GEEBUNG QUEENSLAND. CERTIFICATE NUMBERS: - RN151239, RN151924, RN161940 and CN031-651.

SIGNATURE:

DATE: 25/10/2016

Jeff Kenyon

RE-TEST DUE: October 2017