

Measuretron (Qld) Pty. Limited

A.C.N 010 921 007 A.B.N 23 010 921 007

192 BANCROFT TERRACE DECEPTION BAY QLD 4508

T/A NORTHSIDE / UNIVERSAL SCALES

TEL: 07 3204-0084 FAX: 07 3204-0092

CALIBRATION REPORT

CUSTOMER:

Cathodic Anodes Aust.

8 Endeavour Drive,

KUNDA PARK QLD 4556

CONTACT: Russell Northey

EQUIPMENT:

MAKE NW16104 ID NO. 19
MODEL AS-20/C/W S/N 244359/09
CAPACITY 220g x 0.1mg
LOCATION _____
TRADE APPROVED YES NO

TEST RESULTS

LOAD kg g	AS FOUND kg g	FINAL TEST kg g
0	0.000	0.000
20	19.997 $\bar{5}$	20.000
50	49.996 $\bar{4}$	50.000
100	99.995 $\bar{3}$	100.000
200	199.993 $\bar{6}$	200.000

COMMENTS: Asio Cal Done

ALL TESTING CARRIED OUT USING STANDARD MASSES COVERED BY ONE OR MORE OF THE FOLLOWING CERTIFICATES ISSUED TO MEASURETRON BY NATIONAL MEASUREMENT INSTITUTE, BRISBANE REGIONAL OFFICE, MEASUREMENT STANDARDS LABORATORY, GEEBUNG QUEENSLAND. CERTIFICATE NUMBERS: - RN141872, RN151239, RN151924 and CN031-651.

SIGNATURE:

DATE: 30.09.2015

Jeff Kenyon

RE-TEST DUE: September 2016