

# Measurement Equipment Report

## Multimeters



ID No: 22

Brand: DICK SMITH

Model Q1419

Serial No: 1805 57275

Test Frequency 12 MONTH

### Closed Circuit Potential Test

Test Date	Master	This Meter	Dev.	OK?
11/11/16	-1.063	-1.06	0.003	YES

### Voltage Test 2.77V

Test Date	Master	This Meter	Dev.	OK?
11/11/16	2.64	2.60	0.04	YES

### Voltage Test 12V

Test Date	Master	This Meter	Dev.	OK?
11/11/16	14.8	15.0	0.2	YES

Tested By: PN

Date: 11/11/2016