

Measurement Equipment Report

Multimeters



ID No: 22

Brand: DICK SMITH

Model Q1419.

Serial No: 180557275

Test Frequency 12 MONTH.

Closed Circuit Potential Test

Test Date	Master	This Meter	Dev.	OK?
17/11/15	-1.072	-1.06	-0.012	YES

Voltage Test 2.77V

Test Date	Master	This Meter	Dev.	OK?
19/11/15	2.60	2.50	-1	YES

Voltage Test 12V

Test Date	Master	This Meter	Dev.	OK?
17/11/15	15.3	15.2	-1	YES

Tested By: P.N. Date: 15/11/2015