

Measurement Equipment Report

Multimeters



ID No: 23

Brand: DIGITECH

Model: QM1535

Serial No: 795773

Test Frequency: 12 MONTH

Closed Circuit Potential Test

Test Date	Master	This Meter	Dev.	OK?
11/11/16	-1.063	-1.06	0.003	YES

Voltage Test 2.77V

Test Date	Master	This Meter	Dev.	OK?
11/11/16	2.64	2.60	0.04	YES

Voltage Test 12V

Test Date	Master	This Meter	Dev.	OK?
11/11/16	14.8	14.9	0.1	YES

Tested By: RN

Date: 11/11/2016