

Measurement Equipment Report

Multimeters



ID No: 23

Brand: DCA TECH.

Model QM 1535

Serial No: 795773

Test Frequency 12 MONTH.

Closed Circuit Potential Test

Test Date	Master	This Meter	Dev.	OK?
17/11/15	-1.072	-1.07	-002	YES

Voltage Test 2.77V

Test Date	Master	This Meter	Dev.	OK?
17/11/15	2.60	2.55	-05	YES

Voltage Test 12V

Test Date	Master	This Meter	Dev.	OK?
17/11/15	15.3	15.1	-2	YES

Tested By: RN.

Date: 17/11/2015