

Measurement Equipment Report

Thermometers



All thermometers are calibrated against the MASTER thermometer. For more information refer to procedure – Measurement Equipment

Note: The master thermometer shall only be used for calibration purposes.

Tolerances: Acceptable Temperature Range +/- 20 C from Master.

	Master	General Use			General Use			General Use		
		Temp	Init	OK	Temp	Init	OK	Temp	Init	OK
Brand/Model	TFC-305P	TFC-305P			TFC-305P			TFC-305P		
I.D. No:	27	28			29					
Serial No:	June 2013	June 2013			June 2013					
Frequency:	36 Monthly	12 Monthly			12 Monthly					
Location:	Operations Mgr	Foundry			Foundry					
Date:	Temp – 3 Dips	Temp	Init	OK	Temp	Init	OK	Temp	Init	OK
06 JUNE 2013	506, 506, 508	508	AN	Y	510	AN	OK			
09 Nov 2013	522, 525, 525	528	AN	Y	530	AN	OK			
19 Nov 2014	430, 432, 433	435	AN	Y	432	AN	OK			
12 Nov 2015	504, 504, 503	507	AN	Y	503	AN	OK			
11 Nov 2016	520, 522, 522	525	AN	Y	528	AN	OK			