

Measurement Equipment Report

Multimeters



ID No: 1

Brand: Fluke

Model 75

Serial No: 748 11094

Test Frequency 12 Months

Closed Circuit Potential Test

Test Date	Master	This Meter	Dev.	OK?
<u>N/A</u>	<u>-</u>	<u>-</u>	<u>-</u>	<u>-</u>

Voltage Test 2.77V

Test Date	Master	This Meter	Dev.	OK?
<u>27/8/21</u>	<u>2.782</u>	<u>2.79</u>	<u>0.008</u>	<u>✓</u>

Voltage Test 12V

Test Date	Master	This Meter	Dev.	OK?
<u>27/8/21</u>	<u>12.22</u>	<u>12.21</u>	<u>0.01</u>	<u>✓</u>

Tested By: SP

Date: 27/8/21