

Measurement Equipment Report

Multimeters



ID No: 21

Brand: Riskarb

Model RIS MULTI 145

Serial No: 090 8891

Test Frequency 48 Months

Closed Circuit Potential Test

Test Date	Master	This Meter	Dev.	OK?
n/a	—	—	—	—

Voltage Test 2.77V

Test Date	Master	This Meter	Dev.	OK?
27/8/21		2.762	0.012	✓

Voltage Test 12V

Test Date	Master	This Meter	Dev.	OK?
27/8/21		12.22	0.22	✓

Tested By: SP

Date: 27/8/21