

# Measurement Equipment Report

## Multimeters



ID No: 22

Brand: Dick Smith

Model Q-1419

Serial No: 18055 X275

Test Frequency 12 month.

### Closed Circuit Potential Test

Test Date	Master	This Meter	Dev.	OK?
N/A	-	-	-	-

### Voltage Test 2.77V

Test Date	Master	This Meter	Dev.	OK?
27/8/21	2.762	2.78	0.002	✓

### Voltage Test 12V

Test Date	Master	This Meter	Dev.	OK?
27/8/21	12.22	12.2	0.02	✓

Tested By: SP

Date: 27/8/21