

Measurement Equipment Report  
Multimeters



ID No: 23

Brand: Digitech

Model QM 1535

Serial No: 795773

Test Frequency 12 Month

Closed Circuit Potential Test

Test Date	Master	This Meter	Dev.	OK?

Voltage Test 2.77V

Test Date	Master	This Meter	Dev.	OK?
27/8/21	2.782	2.78	0.002	✓

Voltage Test 12V

Test Date	Master	This Meter	Dev.	OK?
27/8/21	12.22	12.20	0.02	✓

Tested By: SP

Date: 27/8/21