

Measurement Equipment Report

Pyrometers



All Pyrometers are calibrated against the MASTER Pyrometer.

For More Information refer to the procedure – Measurement Equipment.

Note: The Master Pyrometer shall only be used for Calibration purposes.

Tolerances: Acceptable Temperature Range +/- 20 Deg C from Master.

Master	Brand / Model	Frequency	Location
I.D. No: 27	TFC-305P	36 Monthly	Op's Manager
Date:	Dip 1: 515	Dip 2: 580	Dip 3: 493

General Use	Brand / Model	Frequency	Location
I.D. No: 28	TFC-305P	12 Monthly	Foundry
Date:	Dip 1: 508	Dip 2: 585	Dip 3: 493

General Use	Brand / Model	Frequency	Location
I.D. No: 29	TFC-305P	12 Monthly	Foundry
Date:	Dip 1: 510	Dip 2: 584	Dip 3: 496

Tested By:

Date: