

Measuretron (Qld) Pty. Limited

A.C.N 010 921 007 A.B.N 23 010 921 007

192 BANCROFT TERRACE DECEPTION BAY QLD 4508

T/A NORTHSIDE / UNIVERSAL SCALES

TEL: 07 3204-0084 FAX: 07 3204-0092

CALIBRATION REPORT

CUSTOMER:

EQUIPMENT:

Cathodic Anodes Aust.

8 Endeavour Drive,

KUNDA PARK QLD 4556

Make: Ishida MB-300E

Cap: 3000g x 0.1g

S/N: 05340047 I/D: 04

TRADE APPROVED YES NO

CONTACT: Sam Pohlman

TEST RESULTS

LOAD kg g	AS FOUND kg g	FINAL TEST kg g
0	0.0	0.0
200	200.0	200.0
500	500.0	500.0
1000	1000.0	1000.0
2000	2000.0	2000.0
3000	3000.0	3000.0

COMMENTS:

ALL TESTING CARRIED OUT USING STANDARD MASSES COVERED BY ONE OR MORE OF THE FOLLOWING CERTIFICATES ISSUED TO MEASURETRON BY NATIONAL MEASUREMENT INSTITUTE, BRISBANE REGIONAL OFFICE, MEASUREMENT STANDARDS LABORATORY, GEEBUNG QUEENSLAND. CERTIFICATE NUMBERS: - RN201575, CN127453, RN211518 and RN211519.

SIGNATURE:

DATE: 01.10.2021

Jeff Kenyon

RE-TEST DUE: September 2022