

# Measuretron (Qld) Pty. Limited

A.C.N 010 921 007 A.B.N 23 010 921 007

192 BANCROFT TERRACE DECEPTION BAY QLD 4508

T/A NORTHSIDE / UNIVERSAL SCALES

TEL: 07 3204-0084 FAX: 07 3204-0092

## CALIBRATION REPORT

CUSTOMER:

Cathodic Anodes Aust.

8 Endeavour Drive,

KUNDA PARK QLD 4556

CONTACT: Sam Pohlman

EQUIPMENT:

Make: Nuweigh X320/IH1949

Cap: 3000kg x 1kg

S/N: 3506726/100184 I/D:

TRADE APPROVED YES NO

## TEST RESULTS

LOAD kg $\neq$	AS FOUND kg $\neq$	FINAL TEST kg $\neq$
0	0.0	0.0
500	500	500
1000	1000	1000

COMMENTS:

ALL TESTING CARRIED OUT USING STANDARD MASSES COVERED BY ONE OR MORE OF THE FOLLOWING CERTIFICATES ISSUED TO MEASURETRON BY NATIONAL MEASUREMENT INSTITUTE, BRISBANE REGIONAL OFFICE, MEASUREMENT STANDARDS LABORATORY, GEEBUNG QUEENSLAND. CERTIFICATE NUMBERS: - RN201575, CN127453, RN211518 and RN211519.

SIGNATURE:

DATE: 01.10.2021

*Jeff Kenyon*

RE-TEST DUE: September 2022