

# Measuretron (Qld) Pty. Limited

A.C.N 010 921 007 A.B.N 23 010 921 007

192 BANCROFT TERRACE DECEPTION BAY QLD 4508

T/A NORTHSIDE / UNIVERSAL SCALES

TEL: 07 3204-0084 FAX: 07 3204-0092

## CALIBRATION REPORT

CUSTOMER:

EQUIPMENT:

Cathodic Anodes Aust.

8 Endeavour Drive,

KUNDA PARK QLD 4556

Make: Rinstrum R321

Cap: 250kg x 0.1kg

S/N: 3111429 I/D: 07

TRADE APPROVED YES NO

CONTACT: Sam Pohlman

### TEST RESULTS

LOAD kg <i>g</i>	AS FOUND kg <i>g</i>	FINAL TEST kg <i>g</i>
0	0.0	0.0
20	20.0	20.0
40	40.0	40.0
80	80.0	80.0
120	120.0	120.0
150	150.0	150.0

COMMENTS: AFTER ADJUSTMENT OF BASEWORK STOK.

ALL TESTING CARRIED OUT USING STANDARD MASSES COVERED BY ONE OR MORE OF THE FOLLOWING CERTIFICATES ISSUED TO MEASURETRON BY NATIONAL MEASUREMENT INSTITUTE, BRISBANE REGIONAL OFFICE, MEASUREMENT STANDARDS LABORATORY, GEEBUNG QUEENSLAND. CERTIFICATE NUMBERS: - RN201575, CN127453, RN211518 and RN211519.

SIGNATURE:

DATE: 01.10.2021

*Jeff Kenyon*

RE-TEST DUE: September 2022