

Measuretron (Qld) Pty. Limited

A.C.N 010 921 007 A.B.N 23 010 921 007

192 BANCROFT TERRACE DECEPTION BAY QLD 4508

T/A NORTHSIDE / UNIVERSAL SCALES

TEL: 07 3204-0084 FAX: 07 3204-0092

CALIBRATION REPORT

CUSTOMER:

EQUIPMENT:

Cathodic Anodes Aust.

Make: Nuweigh JAD525

8 Endeavour Drive,

Cap: 15kg x 0.005kg

S/N: 706187 I/D: 12

KUNDA PARK QLD 4556

TRADE APPROVED YES NO

CONTACT: Sam Pohlman

TEST RESULTS

LOAD kg g	AS FOUND kg g	FINAL TEST kg g
0	0.000	0.000
1	1.000	1.000
2	2.998	2.000
5	4.996	5.000
8		8.000
10		10.000
15		15.000

COMMENTS:

ALL TESTING CARRIED OUT USING STANDARD MASSES COVERED BY ONE OR MORE OF THE FOLLOWING CERTIFICATES ISSUED TO MEASURETRON BY NATIONAL MEASUREMENT INSTITUTE, BRISBANE REGIONAL OFFICE, MEASUREMENT STANDARDS LABORATORY, GEEBUNG QUEENSLAND. CERTIFICATE NUMBERS: - RN201575, CN127453, RN211518 and RN211519.

SIGNATURE:

DATE: 01.10.2021

Jeff Kenyon

RE-TEST DUE: September 2022