

# Measuretron (Qld) Pty. Limited

A.C.N 010 921 007 A.B.N 23 010 921 007

192 BANCROFT TERRACE DECEPTION BAY QLD 4508

T/A NORTHSIDE / UNIVERSAL SCALES

TEL: 07 3204-0084 FAX: 07 3204-0092

## CALIBRATION REPORT

CUSTOMER:

Cathodic Anodes Aust.

8 Endeavour Drive,

KUNDA PARK QLD 4556

CONTACT: Sam Pohlman

EQUIPMENT:

Make: Nuweigh AS-220

Cap: 220g x 0.001g

S/N: 244359/09 I/D: 19

TRADE APPROVED YES NO

## TEST RESULTS

LOAD $\frac{1}{2}$ g g	AS FOUND $\frac{1}{2}$ g g	FINAL TEST $\frac{1}{2}$ g g
0	0.000	0.000
20	20.000	20.000
50	50.000	50.000
100	100.000	100.000
200	200.000	200.000

COMMENTS:

ALL TESTING CARRIED OUT USING STANDARD MASSES COVERED BY ONE OR MORE OF THE FOLLOWING CERTIFICATES ISSUED TO MEASURETRON BY NATIONAL MEASUREMENT INSTITUTE, BRISBANE REGIONAL OFFICE, MEASUREMENT STANDARDS LABORATORY, GEEBUNG QUEENSLAND. CERTIFICATE NUMBERS: - RN201575, CN127453, RN211518 and RN211519.

SIGNATURE:

DATE: 01.10.2021

*Jeff Kenyon*

RE-TEST DUE: September 2022