

Measurement Equipment Report

Pyrometers



All Pyrometers are calibrated against the MASTER Pyrometer. For More Information refer to the procedure -- Measurement Equipment. Note: The Master Pyrometer shall only be used for Calibration purposes. Tolerances: Acceptable Temperature Range +/- 20 Deg C from Master.

Master	Brand/Model	Frequency	Location
I.D No: 27	TFC-305P	36 Monthly	Op's Manager
Date:	Dip 1 503	Dip 2 503	Dip 3 490

General Use	Brand/Model	Frequency	Location
I.D No: 28	TFC-305P	12 Monthly	Foundry
Date:	Dip 1 493	Dip 2 492	Dip 3 488

General Use	Brand/Model	Frequency	Location
I.D No: 29	TFC-305P	12 Monthly	Foundry
Date:	Dip 1 492	Dip 2 492	Dip 3 489.

Tested by: Sam Pahlman
Duncan Ferguson

Date: 13/10/2022.